

Contract Number: 800439

Certificate: SGSNA/18/SUW/00233
Certificate Project Number: SUW-CERT181100304

Certified Product: Accelerometers
Trademarks (if any): STI
Model(s): CMCP700 Series, CMCP422 Series



Technical Data: $U_i = 28V, I_i = 93mA, P_i = 0.65W; U_i = 28V, I_i = 115mA, P_i = 0.65W; U_i = 44V, I_i = 117mA, P_i = 0.722W$; Class I, II, III, Division 1, 2, Groups A - G, T4, -55°C to +110°C, IP65; Class I, Zone 0, AEx, ia, IIC, T4, Ga, -55°C to +110°C; Zone 20, AEx, ia, IIC, T130°C, IP65, Da, -55°C to +110°C; Class I, II, III, Division 1, 2, Groups A - G, T6, -55°C to +60°C; Zone 20, AEx, ia, IIC, T80°C, IP65, DA, -55°C to +60°C; Class I, II, III, Division 1, 2, Groups A - G, T4, -40°C to +110°C; Class I, Zone 0, AEx, ia, IIC, T4, Ga, -40°C to +110°C; Zone 20, AEx, ia, IIC, T130°C, IP65, DA, -40°C to +110°C; Class I, II, III, Division 1, 2, Groups A - G, T6, -40°C to +60°C; Class I, Zone 0, AEx, ia, IIC, T6, Ga, -40°C to +60°C; Zone 20, AEx, ia, IIC, T80°C, IP65, DA, -40°C to +60°C

Name and Address of Certificate Holder: STI Vibration Monitoring Inc.
1010 East Main Street, 77573, League City, Texas,
United States of America

Name and Address of Manufacturer: STI Vibration Monitoring Inc.
1010 East Main Street, 77573, League City, Texas,
United States of America


This certificate supercedes previous certificates issued with the same certificate number. Certification is valid when products are indicated on the SGS directory of certified products at www.sgs.com. The product is certified according to ISO/IEC Guide 17067, *Conformity assessment - Fundamentals of product certification, System 3*, and in accordance with

UL 913, 8th Ed. Rev. December 6, 2013
CAN/CSA C22.2 No. 157-92 (R2012) +Upd1 +Upd2

Authorized by Effective date: 27 November 2018


Paul Krauss
Certifier

Consumer and Retail Services, a division of SGS North America Inc.
620 Old Peachtree Road, Ste. 100, Suwanee, GA 30024, USA
t +1 770 570 1800 f +1 770 277 1240 www.sgs.com
Page 1 of 2



SGS

Contract Number: 800439

Certificate: SGSNA/18/SUW/00233
Certificate Project Number: SUW-CERT181100304

Certified Product: Accelerometers

Trademarks (if any): STI

Model(s): CMCP700 Series, CMCP422 Series

Name of Certificate Holder: STI Vibration Monitoring Inc.

Name and Address of Factory: STI Vibration Monitoring Inc.
1010 East Main Street, 77573, League City, Texas,
United States of America

Effective date: 27 November 2018

Consumer and Retail Services, a division of SGS North America Inc.
620 Old Peachtree Road, Ste. 100, Suwanee, GA 30024, USA
t +1 770 570 1800 f +1 770 277 1240 www.sgs.com
Page 2 of 2



This certificate is issued by the company under its General Conditions for Certification Services accessible at www.sgs.com/terms_and_conditions.htm. Attention is drawn to the limitations of liability defined therein and in the Test Report here above mentioned which findings are reflected in this Certificate. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.